

Dräger safety

Alcotest<sup>®</sup> 7110 MKIII-C

## CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MKIII-C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MKIII-C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification Date:


02/21/2007

SERIAL NUMBER:

ARXB - 0076

Draeger Safety Diagnostics, Inc.  
Durango, CO

CRD



**Dräger**

## **CERTIFICATE OF ACCURACY**

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.  
(F.R. Vol. 59 No. 249 12/19/94 Notices)

**Draeger Safety Diagnostics, Inc.**

☒ Model: ALCOTEST® CU34

☐ Model: MARK IIA

☐ Other: \_\_\_\_\_

Serial Number:

DD 1B53-0001

Certification Date

Technician

Re-Certification Due Date

7-30-12

JS

7-30-13

**Dräger**

**ALCOTEST® 7110 TEMPERATURE PROBE**

## **CERTIFICATE OF ACCURACY**

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

Certification date:

Next Certification due:

DDXAP2-118

7-30-12

7-30-13

Probe Value

102

**Draeger Safety Diagnostics, Inc.**  
Technical Service Department

JS