

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00136	Calib. Date:	08/22/2012
Certification File No.:	00137	Cert. Date:	08/22/2012
Linearity File No.:	00138	Lin. Date:	08/22/2012
Solution File No.:	00143	Soln. Date:	10/26/2012
Sequential File No.:	00143	File Date:	10/26/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	11K096	Expires:	11/23/2013
		Bottle No.:	0984

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:01D		
Control 1 EC	0.100%	21:02D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03D		
Control 2 EC	0.098%	21:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04D		
Control 3 EC	0.098%	21:05D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	21:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 10/26/2012