

# Calibrating Unit

## New Standard Solution Report

|                         |                       |              |              |
|-------------------------|-----------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.:  | ARXB-0076    |
| Location:               | SPRING LAKE POLICE    |              |              |
| Calibration File No.:   | 00136                 | Calib. Date: | 08/22/2012   |
| Certification File No.: | 00137                 | Cert. Date:  | 08/22/2012   |
| Linearity File No.:     | 00138                 | Lin. Date:   | 08/22/2012   |
| Solution File No.:      | 00145                 | Soln. Date:  | 11/30/2012   |
| Sequential File No.:    | 00145                 | File Date:   | 11/30/2012   |
| Calibrating Unit:       | WET                   | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                | Serial No.:  | DDYB S3-0001 |
| Solution Control Lot:   | 11K096                | Expires:     | 11/23/2013   |
|                         |                       | Bottle No.:  | 0995         |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 12:00S        |                               |                           |
| Control 1 EC      | 0.103%         | 12:01S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.101%         | 12:01S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 12:01S        |                               |                           |
| Control 2 EC      | 0.100%         | 12:02S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.101%         | 12:02S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 12:03S        |                               |                           |
| Control 3 EC      | 0.100%         | 12:03S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.101%         | 12:03S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 12:04S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KUCINSKI

First Name: PTL CHRIS

MI: J

Signature: \_\_\_\_\_

*PTL Chris Kucinski*

Badge No.: 118

Date: 11/30/2012