

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00122	Calib. Date:	02/28/2012
Certification File No.:	00123	Cert. Date:	02/28/2012
Linearity File No.:	00124	Lin. Date:	02/28/2012
Solution File No.:	00134	Soln. Date:	07/01/2012
Sequential File No.:	00134	File Date:	07/01/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	11K096	Expires:	11/23/2013
		Bottle No.:	0991

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:28D		
Control 1 EC	0.100%	16:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:29D		
Control 2 EC	0.097%	16:30D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:30D		
Control 3 EC	0.098%	16:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: Sgt P. Gilroy #115

Badge No.: 115

Date: 07/01/2012