

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00136 Calib. Date: 08/22/2012 Calib. No.: 00010
Certification File No.: 00137 Cert. Date: 08/22/2012 Cert. No.: 00009
Linearity File No.: 00138 Lin. Date: 08/22/2012 Lin. No.: 00009
Solution File No.: 00152 Soln. Date: 01/20/2013 Soln. No.: 00065
Sequential File No.: 00152 File Date: 01/20/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001
Control Solution %: 0.100% Expires: 11/23/2013
Solution Control Lot: 11K096 Bottle No.: 0979

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:04S		
Control 1 EC	0.101%	12:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05S		
Control 2 EC	0.098%	12:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:07S		
Control 3 EC	0.099%	12:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ZOING

First Name: ROBERT

MI: -

Signature:  #111

Badge No.: 111

Date: 01/20/2013