

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00136 Calib. Date: 08/22/2012 Calib. No.: 00010
Certification File No.: 00137 Cert. Date: 08/22/2012 Cert. No.: 00009
Linearity File No.: 00138 Lin. Date: 08/22/2012 Lin. No.: 00009
Solution File No.: 00154 Soln. Date: 01/20/2013 Soln. No.: 00066
Sequential File No.: 00154 File Date: 01/20/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001
Control Solution %: 0.100% Expires: 11/23/2013
Solution Control Lot: 11K096 Bottle No.: 0989

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:47S		
Control 1 EC	0.101%	23:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:49S		
Control 2 EC	0.099%	23:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:50S		
Control 3 EC	0.099%	23:51S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:52S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature:  115

Badge No.: 115

Date: 01/20/2013