

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00156	Calib. Date:	02/21/2013
Certification File No.:	00157	Cert. Date:	02/21/2013
Linearity File No.:	00158	Lin. Date:	02/21/2013
Solution File No.:	00162	Soln. Date:	04/14/2013
Sequential File No.:	00162	File Date:	04/14/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	11K096	Expires:	11/23/2013
		Bottle No.:	0978

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:49D		
Control 1 EC	0.098%	11:49D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:50D		
Control 2 EC	0.096%	11:51D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52D		
Control 3 EC	0.096%	11:52D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ZOINO

First Name: ROBERT

MI: -

Signature: _____

Badge No.: 111

Date: 04/14/2013