

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00156 Calib. Date: 02/21/2013 Calib. No.: 00011
Certification File No.: 00157 Cert. Date: 02/21/2013 Cert. No.: 00010
Linearity File No.: 00158 Lin. Date: 02/21/2013 Lin. No.: 00010
Solution File No.: 00163 Soln. Date: 05/12/2013 Soln. No.: 00071
Sequential File No.: 00163 File Date: 05/12/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001
Control Solution %: 0.100% Expires: 11/23/2013
Solution Control Lot: 11K096 Bottle No.: 0987

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:31D		
Control 1 EC	0.098%	02:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:33D		
Control 2 EC	0.096%	02:33D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	02:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:34D		
Control 3 EC	0.096%	02:35D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KUCINSKI

First Name: CHRIS

MI: J

Signature: Det. C. K. #118

Badge No.: 118

Date: 05/12/2013