

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00168 Calib. Date: 07/23/2013 Calib. No.: 00012
Certification File No.: 00169 Cert. Date: 07/23/2013 Cert. No.: 00011
Linearity File No.: 00170 Lin. Date: 07/23/2013 Lin. No.: 00011
Solution File No.: 00183 Soln. Date: 01/03/2014 Soln. No.: 00081
Sequential File No.: 00183 File Date: 01/03/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0002
Control Solution %: 0.100% Expires: 06/17/2015
Solution Control Lot: 13F114 Bottle No.: 0701

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:47S		
Control 1 EC	0.099%	21:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:49S		
Control 2 EC	0.098%	21:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:50S		
Control 3 EC	0.097%	21:51S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:52S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PLOSKONKA

First Name: ANTHONY

MI: M

Badge No.: 113

Date: 01/03/2014

Signature: 