

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00184 Calib. Date: 01/10/2014 Calib. No.: 00013
Certification File No.: 00185 Cert. Date: 01/10/2014 Cert. No.: 00012
Linearity File No.: 00186 Lin. Date: 01/10/2014 Lin. No.: 00012
Solution File No.: 00188 Soln. Date: 02/02/2014 Soln. No.: 00083
Sequential File No.: 00188 File Date: 02/02/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001
Control Solution %: 0.100% Expires: 06/17/2015
Solution Control Lot: 13F114 Bottle No.: 0710

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:54S		
Control 1 EC	0.098%	18:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	18:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:56S		
Control 2 EC	0.095%	18:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	18:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:58S		
Control 3 EC	0.096%	18:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	18:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

Patrick Gilroy

Badge No.: 115

Date: 02/02/2014