

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00184	Calib. Date:	01/10/2014
Certification File No.:	00185	Cert. Date:	01/10/2014
Linearity File No.:	00186	Lin. Date:	01/10/2014
Solution File No.:	00191	Soln. Date:	03/23/2014
Sequential File No.:	00191	File Date:	03/23/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	13F114	Expires:	06/17/2015
		Bottle No.:	0702

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:42S		
Control 1 EC	0.100%	20:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:44S		
Control 2 EC	0.096%	20:44S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:45S		
Control 3 EC	0.096%	20:46S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

Pat P. Gilroy

Badge No.: 115

Date: 03/23/2014