

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00184	Calib. Date:	01/10/2014
Certification File No.:	00185	Cert. Date:	01/10/2014
Linearity File No.:	00186	Lin. Date:	01/10/2014
Solution File No.:	00193	Soln. Date:	04/19/2014
Sequential File No.:	00193	File Date:	04/19/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	13F114	Expires:	06/17/2015
		Bottle No.:	0703

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:55D		
Control 1 EC	0.100%	21:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:57D		
Control 2 EC	0.097%	21:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:58D		
Control 3 EC	0.098%	21:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:00D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PLOSKONKA

First Name: ANTHONY

MI: M

Signature: 

Badge No.: 113

Date: 04/19/2014