

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00168	Calib. Date:	07/23/2013
Certification File No.:	00169	Cert. Date:	07/23/2013
Linearity File No.:	00170	Lin. Date:	07/23/2013
Solution File No.:	00174	Soln. Date:	09/12/2013
Sequential File No.:	00174	File Date:	09/12/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	11K096	Expires:	11/23/2013
		Bottle No.:	0980

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:09D		
Control 1 EC	0.100%	10:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:10D		
Control 2 EC	0.098%	10:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12D		
Control 3 EC	0.098%	10:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____



Badge No.: 115

Date: 09/12/2013