

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00237	Calib. Date:	07/10/2015
Certification File No.:	00238	Cert. No.:	00017
Linearity File No.:	00239	Cert. Date:	07/10/2015
Solution File No.:	00249	Lin. Date:	07/10/2015
Sequential File No.:	00249	Lin. No.:	00015
		Soln. Date:	11/04/2015
		Soln. No.:	00111
		File Date:	11/04/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	14L135	Expires:	12/31/2016
		Bottle No.:	0800

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:28S		
Control 1 EC	0.101%	00:29S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	00:29S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:29S		
Control 2 EC	0.100%	00:30S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	00:30S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:31S		
Control 3 EC	0.100%	00:31S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	00:31S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

Pat Gilroy

Badge No.: 115

Date: 11/04/2015