

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXB-0076
Location:	SPRING LAKE POLICE	
Calibration File No.:	00237	Calib. Date: 07/10/2015
Certification File No.:	00238	Calib. No.: 00017
Linearity File No.:	00239	Cert. Date: 07/10/2015
Solution File No.:	00255	Cert. No.: 00015
Sequential File No.:	00255	Lin. Date: 07/10/2015
		Lin. No.: 00015
		Soln. Date: 12/15/2015
		Soln. No.: 00113
		File Date: 12/15/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDYB S3-0002
Solution Control Lot:	14L135	Expires: 12/31/2016
		Bottle No.: 0783

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:01S		
Control 1 EC	0.101%	02:01S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	02:01S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:02S		
Control 2 EC	0.100%	02:03S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	02:03S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:03S		
Control 3 EC	0.100%	02:04S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:04S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 12/15/2015