

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00262 Calib. Date: 01/08/2016 Calib. No.: 00018
Certification File No.: 00263 Cert. Date: 01/08/2016 Cert. No.: 00016
Linearity File No.: 00264 Lin. Date: 01/08/2016 Lin. No.: 00016
Solution File No.: 00269 Soln. Date: 02/15/2016 Soln. No.: 00116
Sequential File No.: 00269 File Date: 02/15/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001
Control Solution %: 0.100% Expires: 12/31/2016
Solution Control Lot: 14L135 Bottle No.: 0792

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:44S		
Control 1 EC	0.100%	10:45S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45S		
Control 2 EC	0.099%	10:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:47S		
Control 3 EC	0.099%	10:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:48S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: SA P. Gilroy

Badge No.: 115

Date: 02/15/2016