

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00262	Calib. Date:	01/08/2016
Certification File No.:	00263	Cert. Date:	01/08/2016
Linearity File No.:	00264	Lin. Date:	01/08/2016
Solution File No.:	00274	Soln. Date:	04/10/2016
Sequential File No.:	00274	File Date:	04/10/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	14L135	Expires:	12/31/2016
		Bottle No.:	0781

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:35D		
Control 1 EC	0.099%	10:36D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:36D		
Control 2 EC	0.099%	10:37D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:38D		
Control 3 EC	0.099%	10:38D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____



Badge No.: 115

Date: 04/10/2016