

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00282	Calib. Date:	06/06/2016
Certification File No.:	00283	Cert. Date:	06/06/2016
Linearity File No.:	00284	Lin. Date:	06/06/2016
Solution File No.:	00289	Soln. Date:	06/25/2016
Sequential File No.:	00289	File Date:	06/25/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	201603L	Expires:	03/22/2018
		Bottle No.:	1160

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:23D		
Control 1 EC	0.099%	19:24D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:25D		
Control 2 EC	0.098%	19:25D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:26D		
Control 3 EC	0.099%	19:27D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature:

Pat Gilroy #115

Badge No.: 115

Date: 06/25/2016