

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SPRING LAKE POLICE
Calibration File No.: 00282
Certification File No.: 00283
Linearity File No.: 00284
Solution File No.: 00295
Sequential File No.: 00295

Serial No.: ARXB-0076

Calib. Date: 06/06/2016
Cert. Date: 06/06/2016
Lin. Date: 06/06/2016
Soln. Date: 07/18/2016
File Date: 07/18/2016

Calib. No.: 00019
Cert. No.: 00017
Lin. No.: 00017
Soln. No.: 00125

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 201603L

Model No.: CU-34

Serial No.: DDYB S3-0002
Expires: 03/22/2018
Bottle No.: 1162

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:32D		
Control 1 EC	0.101%	12:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:33D		
Control 2 EC	0.099%	12:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:35D		
Control 3 EC	0.099%	12:35D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:35D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

Pat Gilroy #115

Badge No.: 115

Date: 07/18/2016