

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00282 Calib. Date: 06/06/2016 Calib. No.: 00019
Certification File No.: 00283 Cert. Date: 06/06/2016 Cert. No.: 00017
Linearity File No.: 00284 Lin. Date: 06/06/2016 Lin. No.: 00017
Solution File No.: 00299 Soln. Date: 08/07/2016 Soln. No.: 00126
Sequential File No.: 00299 File Date: 08/07/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0002
Control Solution %: 0.100% Expires: 03/22/2018
Solution Control Lot: 201603L Bottle No.: 1254

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:04D		
Control 1 EC	0.098%	22:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:05D		
Control 2 EC	0.097%	22:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:07D		
Control 3 EC	0.097%	22:07D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: Pat P. Gilroy 115

Badge No.: 115

Date: 08/07/2016