

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SPRING LAKE POLICE
Calibration File No.: 00282
Certification File No.: 00283
Linearity File No.: 00284
Solution File No.: 00332
Sequential File No.: 00332

Serial No.: ARXB-0076
Calib. Date: 06/06/2016
Cert. Date: 06/06/2016
Lin. Date: 06/06/2016
Soln. Date: 09/26/2016
File Date: 09/26/2016
Calib. No.: 00019
Cert. No.: 00017
Lin. No.: 00017
Soln. No.: 00129

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 201603L

Model No.: CU-34
Serial No.: DDYB S3-0002
Expires: 03/22/2018
Bottle No.: 1251

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:26D		
Control 1 EC	0.102%	11:26D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27D		
Control 2 EC	0.100%	11:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28D		
Control 3 EC	0.100%	11:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____



Badge No.: 115

Date: 09/26/2016