

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SPRING LAKE POLICE
Serial No.: ARXB-0076
Calibration File No.: 00339
Calib. Date: 11/30/2016
Calib. No.: 00020
Certification File No.: 00340
Cert. Date: 11/30/2016
Cert. No.: 00018
Linearity File No.: 00341
Lin. Date: 11/30/2016
Lin. No.: 00018
Solution File No.: 00344
Soln. Date: 12/18/2016
Soln. No.: 00134
Sequential File No.: 00344
File Date: 12/18/2016
Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDYB S3-0001
Control Solution %: 0.100%
Expires: 09/07/2018
Solution Control Lot: 201609A
Bottle No.: 0595

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:27S		
Control 1 EC	0.102%	11:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28S		
Control 2 EC	0.101%	11:29S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:29S		
Control 3 EC	0.101%	11:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 12/18/2016