

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00354	Calib. Date:	05/26/2017
Certification File No.:	00355	Cert. Date:	05/26/2017
Linearity File No.:	00356	Lin. Date:	05/26/2017
Solution File No.:	00361	Soln. Date:	07/03/2017
Sequential File No.:	00361	File Date:	07/03/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	201609A	Expires:	09/07/2018
		Bottle No.:	0591

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:47D		
Control 1 EC	0.101%	13:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49D		
Control 2 EC	0.100%	13:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50D		
Control 3 EC	0.100%	13:51D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:51D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 07/03/2017