

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00354 Calib. Date: 05/26/2017 Calib. No.: 00021
Certification File No.: 00355 Cert. Date: 05/26/2017 Cert. No.: 00019
Linearity File No.: 00356 Lin. Date: 05/26/2017 Lin. No.: 00019
Solution File No.: 00366 Soln. Date: 09/04/2017 Soln. No.: 00147
Sequential File No.: 00366 File Date: 09/04/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0002
Control Solution %: 0.100% Expires: 09/07/2018
Solution Control Lot: 201609A Bottle No.: 0590

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:37D		
Control 1 EC	0.101%	22:37D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	22:37D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:38D		
Control 2 EC	0.100%	22:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:39D		
Control 3 EC	0.100%	22:40D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____



Badge No.: 115

Date: 09/04/2017