

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXB-0076
Location:	SPRING LAKE POLICE	
Calibration File No.:	00339	Calib. Date: 11/30/2016
Certification File No.:	00340	Calib. No.: 00020
Linearity File No.:	00341	Cert. Date: 11/30/2016
Solution File No.:	00350	Cert. No.: 00018
Sequential File No.:	00350	Lin. Date: 11/30/2016
		Lin. No.: 00018
		Soln. Date: 04/02/2017
		Soln. No.: 00139
		File Date: 04/02/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDYB S3-0001
Solution Control Lot:	201609A	Expires: 09/07/2018
		Bottle No.: 0583

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:53D		
Control 1 EC	0.101%	22:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	22:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:54D		
Control 2 EC	0.101%	22:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:56D		
Control 3 EC	0.100%	22:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 04/02/2017