

# Calibrating Unit

## New Standard Solution Report

|                         |                      |                         |
|-------------------------|----------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKII-C | Serial No.: ARXB-0076   |
| Location:               | SPRING LAKE POLICE   |                         |
| Calibration File No.:   | 00374                | Calib. Date: 11/17/2017 |
| Certification File No.: | 00375                | Cert. No.: 00022        |
| Linearity File No.:     | 00376                | Cert. Date: 11/17/2017  |
| Solution File No.:      | 00383                | Lin. No.: 00020         |
| Sequential File No.:    | 00383                | Soln. No.: 00153        |
|                         |                      | File Date: 01/21/2018   |

|                       |         |                  |                          |
|-----------------------|---------|------------------|--------------------------|
| Calibrating Unit:     | WET     | Model No.: CU-34 | Serial No.: DDYB S3-0001 |
| Control Solution %:   | 0.100%  |                  | Expires: 09/07/2018      |
| Solution Control Lot: | 201609A |                  | Bottle No.: 0603         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 21:43S |                |                     |
| Control 1 EC      | 0.102% | 21:43S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 21:43S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:44S |                |                     |
| Control 2 EC      | 0.101% | 21:45S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 21:45S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:45S |                |                     |
| Control 3 EC      | 0.100% | 21:46S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 21:46S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 21:47S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature:

*Pat Gilroy*

Badge No.: 115

Date: 01/21/2018