

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00354 **Calib. Date:** 05/26/2017 **Calib. No.:** 00021
Certification File No.: 00355 **Cert. Date:** 05/26/2017 **Cert. No.:** 00019
Linearity File No.: 00356 **Lin. Date:** 05/26/2017 **Lin. No.:** 00019
Solution File No.: 00371 **Soln. Date:** 10/30/2017 **Soln. No.:** 00149
Sequential File No.: 00371 **File Date:** 10/30/2017

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDYB S3-0002
Control Solution %: 0.100% **Expires:** 09/07/2018
Solution Control Lot: 201609A **Bottle No.:** 0587

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:26S		
Control 1 EC	0.100%	20:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	20:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:27S		
Control 2 EC	0.100%	20:28S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:28S		
Control 3 EC	0.100%	20:29S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

Pat Gilroy

Badge No.: 115

Date: 10/30/2017