

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXB-0076
Location:	SPRING LAKE POLICE	
Calibration File No.:	00374	Calib. Date: 11/17/2017
Certification File No.:	00375	Calib. No.: 00022
Linearity File No.:	00376	Cert. Date: 11/17/2017
Solution File No.:	00382	Cert. No.: 00020
Sequential File No.:	00382	Lin. Date: 11/17/2017
		Lin. No.: 00020
		Soln. Date: 12/31/2017
		Soln. No.: 00152
		File Date: 12/31/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDYB S3-0001
Solution Control Lot:	201609A	Expires: 09/07/2018
		Bottle No.: 0602

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:58S		
Control 1 EC	0.100%	15:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:00S		
Control 2 EC	0.100%	16:00S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:01S		
Control 3 EC	0.099%	16:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

Badge No.: 115

Date: 12/31/2017