

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SPRING LAKE POLICE
Calibration File No.: 00401
Certification File No.: 00402
Linearity File No.: 00403
Solution File No.: 00405
Sequential File No.: 00405

Serial No.: ARXB-0076

Calib. Date: 08/17/2018
Cert. Date: 08/17/2018
Lin. Date: 08/17/2018
Soln. Date: 09/09/2018
File Date: 09/09/2018

Calib. No.: 00024
Cert. No.: 00022
Lin. No.: 00022
Soln. No.: 00164

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17110

Model No.: CU-34

Serial No.: DDYB S3-0002
Expires: 03/20/2019
Bottle No.: 0221

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:04D		
Control 1 EC	0.103%	11:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	11:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:06D		
Control 2 EC	0.101%	11:06D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07D		
Control 3 EC	0.101%	11:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____



Badge No.: 115

Date: 09/09/2018