

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00390	Calib. Date:	04/12/2018
Certification File No.:	00391	Cert. Date:	04/12/2018
Linearity File No.:	00392	Lin. Date:	04/12/2018
Solution File No.:	00400	Soln. Date:	07/30/2018
Sequential File No.:	00400	File Date:	07/30/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	17110	Expires:	03/20/2019
		Bottle No.:	0227

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:10D		
Control 1 EC	0.100%	10:11D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:11D		
Control 2 EC	0.099%	10:12D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		
Control 3 EC	0.099%	10:13D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: Pat Gilroy 115

Badge No.: 115

Date: 07/30/2018