

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00374	Calib. Date:	11/17/2017
Certification File No.:	00375	Cert. Date:	11/17/2017
Linearity File No.:	00376	Lin. Date:	11/17/2017
Solution File No.:	00384	Soln. Date:	02/12/2018
Sequential File No.:	00384	File Date:	02/12/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	201609A	Expires:	09/07/2018
		Bottle No.:	0601

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:45S		
Control 1 EC	0.100%	14:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:47S		
Control 2 EC	0.100%	14:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:48S		
Control 3 EC	0.100%	14:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 02/12/2018