

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SPRING LAKE POLICE
Calibration File No.: 00374
Certification File No.: 00375
Linearity File No.: 00376
Solution File No.: 00387
Sequential File No.: 00387

Serial No.: ARXB-0076
Calib. Date: 11/17/2017
Cert. Date: 11/17/2017
Lin. Date: 11/17/2017
Soln. Date: 03/04/2018
File Date: 03/04/2018
Calib. No.: 00022
Cert. No.: 00020
Lin. No.: 00020
Soln. No.: 00155

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 201609A

Model No.: CU-34
Serial No.: DDYB S3-0001
Expires: 09/07/2018
Bottle No.: 0599

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:20S		
Control 1 EC	0.099%	21:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:21S		
Control 2 EC	0.099%	21:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:23S		
Control 3 EC	0.099%	21:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____



Badge No.: 115

Date: 03/04/2018