

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00390 Calib. Date: 04/12/2018 Calib. No.: 00023
Certification File No.: 00391 Cert. Date: 04/12/2018 Cert. No.: 00021
Linearity File No.: 00392 Lin. Date: 04/12/2018 Lin. No.: 00021
Solution File No.: 00397 Soln. Date: 06/17/2018 Soln. No.: 00160
Sequential File No.: 00397 File Date: 06/17/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0002
Control Solution %: 0.100% Expires: 03/20/2019
Solution Control Lot: 17110 Bottle No.: 0621

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:41D		
Control 1 EC	0.098%	10:41D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:42D		
Control 2 EC	0.097%	10:42D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:43D		
Control 3 EC	0.097%	10:44D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	10:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

PA Gilroy

Badge No.: 115

Date: 06/17/2018