

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00401	Calib. Date:	08/17/2018
Certification File No.:	00402	Cert. Date:	08/17/2018
Linearity File No.:	00403	Lin. Date:	08/17/2018
Solution File No.:	00409	Soln. Date:	11/12/2018
Sequential File No.:	00409	File Date:	11/12/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0002
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0073

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:20S		
Control 1 EC	0.103%	00:20S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	00:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:21S		
Control 2 EC	0.102%	00:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	00:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:22S		
Control 3 EC	0.101%	00:23S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	00:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____

Pat Gilroy

Badge No.: 115

Date: 11/12/2018