

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXB-0076
Location:	SPRING LAKE POLICE	
Calibration File No.:	00401	Calib. Date: 08/17/2018
Certification File No.:	00402	Calib. No.: 00024
Linearity File No.:	00403	Cert. Date: 08/17/2018
Solution File No.:	00415	Cert. No.: 00022
Sequential File No.:	00415	Lin. Date: 08/17/2018
		Lin. No.: 00022
		Soln. Date: 01/10/2019
		Soln. No.: 00170
		File Date: 01/10/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDYB S3-0002
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0136

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:10S		
Control 1 EC	0.101%	17:11S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:11S		
Control 2 EC	0.100%	17:12S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:13S		
Control 3 EC	0.100%	17:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	17:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KELLY

First Name: WILLIAM

MI: E

Signature:



Badge No.: 124

Date: 01/10/2019