

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SPRING LAKE POLICE
Calibration File No.: 00416
Certification File No.: 00417
Linearity File No.: 00418
Solution File No.: 00425
Sequential File No.: 00425

Serial No.: ARXB-0076

Calib. Date: 01/24/2019
Cert. Date: 01/24/2019
Lin. Date: 01/24/2019
Soln. Date: 03/06/2019
File Date: 03/06/2019

Calib. No.: 00025
Cert. No.: 00023
Lin. No.: 00023
Soln. No.: 00174

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 18050

Model No.: CU-34

Serial No.: DDYB S3-0001
Expires: 01/31/2020
Bottle No.: 0025

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:17S		
Control 1 EC	0.099%	01:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:19S		
Control 2 EC	0.099%	01:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:20S		
Control 3 EC	0.098%	01:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	01:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: *Pat Gilroy*

Badge No.: 115

Date: 03/06/2019