

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00416	Calib. Date:	01/24/2019
Certification File No.:	00417	Cert. Date:	01/24/2019
Linearity File No.:	00418	Lin. Date:	01/24/2019
Solution File No.:	00423	Soln. Date:	02/11/2019
Sequential File No.:	00423	File Date:	02/11/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	18050	Expires:	01/31/2020
		Bottle No.:	0037

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:08S		
Control 1 EC	0.099%	19:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:10S		
Control 2 EC	0.099%	19:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:11S		
Control 3 EC	0.099%	19:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____



Badge No.: 115

Date: 02/11/2019