

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00416 Calib. Date: 01/24/2019 Calib. No.: 00025
Certification File No.: 00417 Cert. Date: 01/24/2019 Cert. No.: 00023
Linearity File No.: 00418 Lin. Date: 01/24/2019 Lin. No.: 00023
Solution File No.: 00436 Soln. Date: 05/21/2019 Soln. No.: 00178
Sequential File No.: 00436 File Date: 05/21/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001
Control Solution %: 0.100% Expires: 01/31/2020
Solution Control Lot: 18050 Bottle No.: 0187

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:02D		
Control 1 EC	0.099%	11:02D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:03D		
Control 2 EC	0.099%	11:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:04D		
Control 3 EC	0.098%	11:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: _____



Badge No.: 115

Date: 05/21/2019