

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXB-0076
Location:	SPRING LAKE POLICE	
Calibration File No.:	00416	Calib. Date: 01/24/2019
Certification File No.:	00417	Calib. No.: 00025
Linearity File No.:	00418	Cert. Date: 01/24/2019
Solution File No.:	00439	Cert. No.: 00023
Sequential File No.:	00439	Lin. Date: 01/24/2019
		Lin. No.: 00023
		Soln. Date: 06/09/2019
		Soln. No.: 00179
		File Date: 06/09/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDYB S3-0001
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0244

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:14D		
Control 1 EC	0.098%	22:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:15D		
Control 2 EC	0.097%	22:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:17D		
Control 3 EC	0.097%	22:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	22:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: \_\_\_\_\_

*Pat Gilroy*

Badge No.: 115

Date: 06/09/2019