

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00443	Calib. Date:	06/28/2019
Certification File No.:	00444	Cert. Date:	06/28/2019
Linearity File No.:	00445	Lin. Date:	06/28/2019
Solution File No.:	00454	Soln. Date:	07/07/2019
Sequential File No.:	00454	File Date:	07/07/2019
		Calib. No.:	00026
		Cert. No.:	00024
		Lin. No.:	00024
		Soln. No.:	00182
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	19110	Expires:	04/09/2021
		Bottle No.:	0601

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:06D		
Control 1 EC	0.097%	03:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:07D		
Control 2 EC	0.097%	03:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	03:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:08D		
Control 3 EC	0.097%	03:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	03:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: DUGGAN

First Name: SEAN

MI: T

Signature: _____



Badge No.: 125

Date: 07/07/2019