

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXB-0076  
Location: SPRING LAKE POLICE  
Calibration File No.: 00416 Calib. Date: 01/24/2019 Calib. No.: 00025  
Certification File No.: 00417 Cert. Date: 01/24/2019 Cert. No.: 00023  
Linearity File No.: 00418 Lin. Date: 01/24/2019 Lin. No.: 00023  
Solution File No.: 00430 Soln. Date: 05/01/2019 Soln. No.: 00177  
Sequential File No.: 00430 File Date: 05/01/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001  
Control Solution %: 0.100% Expires: 01/31/2020  
Solution Control Lot: 18050 Bottle No.: 0077

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 01:41D        |                               |                           |
| Control 1 EC      | 0.099%         | 01:41D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.099%         | 01:41D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 01:42D        |                               |                           |
| Control 2 EC      | 0.098%         | 01:43D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.098%         | 01:43D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 01:43D        |                               |                           |
| Control 3 EC      | 0.097%         | 01:44D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 01:44D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 01:45D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 05/01/2019