

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00443 Calib. Date: 06/28/2019 Calib. No.: 00026
Certification File No.: 00444 Cert. Date: 06/28/2019 Cert. No.: 00024
Linearity File No.: 00445 Lin. Date: 06/28/2019 Lin. No.: 00024
Solution File No.: 00466 Soln. Date: 08/10/2019 Soln. No.: 00185
Sequential File No.: 00466 File Date: 08/10/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0002
Control Solution %: 0.100% Expires: 04/09/2021
Solution Control Lot: 19110 Bottle No.: 0540

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:46D		
Control 1 EC	0.097%	11:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:47D		
Control 2 EC	0.096%	11:48D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		
Control 3 EC	0.097%	11:49D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:50D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KELLY

First Name: WILLIAM

MI: E

Signature: 

Badge No.: 124

Date: 08/10/2019