

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXB-0076
Location:	SPRING LAKE POLICE	
Calibration File No.:	00443	Calib. Date: 06/28/2019
Certification File No.:	00444	Calib. No.: 00026
Linearity File No.:	00445	Cert. Date: 06/28/2019
Solution File No.:	00474	Cert. No.: 00024
Sequential File No.:	00474	Lin. Date: 06/28/2019
		Lin. No.: 00024
		Soln. Date: 09/01/2019
		Soln. No.: 00186
		File Date: 09/01/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDYB S3-0002
Solution Control Lot:	19110	Expires: 04/09/2021
		Bottle No.: 0590

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:32D		
Control 1 EC	0.097%	03:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:34D		
Control 2 EC	0.096%	03:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:35D		
Control 3 EC	0.096%	03:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:36D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DUGGAN

First Name: SEAN

MI: T

Signature: _____

Badge No.: 125

Date: 09/01/2019