

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00443 **Calib. Date:** 06/28/2019 **Calib. No.:** 00026
Certification File No.: 00444 **Cert. Date:** 06/28/2019 **Cert. No.:** 00024
Linearity File No.: 00445 **Lin. Date:** 06/28/2019 **Lin. No.:** 00024
Solution File No.: 00497 **Soln. Date:** 11/22/2019 **Soln. No.:** 00191
Sequential File No.: 00497 **File Date:** 11/22/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDYB S3-0002
Control Solution %: 0.100% **Expires:** 06/10/2021
Solution Control Lot: 19150 **Bottle No.:** 1206

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:01S		
Control 1 EC	0.098%	02:01S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:02S		
Control 2 EC	0.097%	02:03S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:03S		
Control 3 EC	0.098%	02:04S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	02:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: PHILLIPS

First Name: MATTHEW

MI: J

Signature: Matthew Phillips #126

Badge No.: 126

Date: 11/22/2019