

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SPRING LAKE POLICE
Calibration File No.: 00443
Certification File No.: 00444
Linearity File No.: 00445
Solution File No.: 00486
Sequential File No.: 00486

Serial No.: ARXB-0076

Calib. Date: 06/28/2019
Cert. Date: 06/28/2019
Lin. Date: 06/28/2019
Soln. Date: 10/21/2019
File Date: 10/21/2019

Calib. No.: 00026
Cert. No.: 00024
Lin. No.: 00024
Soln. No.: 00189

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 19110

Model No.: CU-34

Serial No.: DDYB S3-0002
Expires: 04/09/2021
Bottle No.: 0596

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:47D		
Control 1 EC	0.097%	17:48D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:48D		
Control 2 EC	0.097%	17:49D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	17:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:50D		
Control 3 EC	0.097%	17:50D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: Pat Gilroy 115

Badge No.: 115

Date: 10/21/2019