

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00443 Calib. Date: 06/28/2019 Calib. No.: 00026
Certification File No.: 00444 Cert. Date: 06/28/2019 Cert. No.: 00024
Linearity File No.: 00445 Lin. Date: 06/28/2019 Lin. No.: 00024
Solution File No.: 00460 Soln. Date: 07/31/2019 Soln. No.: 00184
Sequential File No.: 00460 File Date: 07/31/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0002
Control Solution %: 0.100% Expires: 04/09/2021
Solution Control Lot: 19110 Bottle No.: 0595

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:00D		
Control 1 EC	0.099%	16:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:01D		
Control 2 EC	0.097%	16:02D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	16:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:02D		
Control 3 EC	0.097%	16:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXAP2-117 (DL)

Changed By:

Last Name: LUTZ

First Name: DENNIS

MI: J.

Signature: Tpc I Dent 7045

Badge No.: 7045

Date: 07/31/2019