

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: SPRING LAKE POLICE
Calibration File No.: 00506
Certification File No.: 00507
Linearity File No.: 00508
Solution File No.: 00520
Sequential File No.: 00520

Serial No.: ARXB-0076

Calib. Date: 12/03/2019
Cert. Date: 12/03/2019
Lin. Date: 12/03/2019
Soln. Date: 01/06/2020
File Date: 01/06/2020

Calib. No.: 00027
Cert. No.: 00025
Lin. No.: 00025
Soln. No.: 00197

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 19110

Model No.: CU-34

Serial No.: DDYB S3-0001
Expires: 04/09/2021
Bottle No.: 0594

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:32S		
Control 1 EC	0.099%	00:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:33S		
Control 2 EC	0.098%	00:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:35S		
Control 3 EC	0.098%	00:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 01/06/2020