

# Calibrating Unit

## New Standard Solution Report

|                         |                       |              |              |
|-------------------------|-----------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.:  | ARXB-0076    |
| Location:               | SPRING LAKE POLICE    |              |              |
| Calibration File No.:   | 00506                 | Calib. Date: | 12/03/2019   |
| Certification File No.: | 00507                 | Cert. Date:  | 12/03/2019   |
| Linearity File No.:     | 00508                 | Lin. Date:   | 12/03/2019   |
| Solution File No.:      | 00515                 | Soln. Date:  | 12/11/2019   |
| Sequential File No.:    | 00515                 | File Date:   | 12/11/2019   |
| Calibrating Unit:       | WET                   | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                | Serial No.:  | DDYB S3-0001 |
| Solution Control Lot:   | 19150                 | Expires:     | 06/10/2021   |
|                         |                       | Bottle No.:  | 0541         |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 01:42S        |                               |                           |
| Control 1 EC      | 0.096%         | 01:43S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.096%         | 01:43S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 01:44S        |                               |                           |
| Control 2 EC      | 0.094%         | 01:44S        | 34.0°C                        | CONTROL TEST FAILED       |
| Control 2 IR      | 0.097%         | 01:44S        | 34.0°C                        | CONTROL TEST FAILED       |
| Ambient Air Blank | 0.000%         | 01:45S        |                               |                           |
| Control 3 EC      | 0.094%         | 01:46S        | 34.0°C                        | CONTROL TEST FAILED       |
| Control 3 IR      | 0.096%         | 01:46S        | 34.0°C                        | CONTROL TEST FAILED       |
| Ambient Air Blank | 0.000%         | 01:47S        |                               |                           |

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 12/11/2019