

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXB-0076
Location: SPRING LAKE POLICE
Calibration File No.: 00506 Calib. Date: 12/03/2019 Calib. No.: 00027
Certification File No.: 00507 Cert. Date: 12/03/2019 Cert. No.: 00025
Linearity File No.: 00508 Lin. Date: 12/03/2019 Lin. No.: 00025
Solution File No.: 00516 Soln. Date: 12/11/2019 Soln. No.: 00195
Sequential File No.: 00516 File Date: 12/11/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDYB S3-0001
Control Solution %: 0.100% Expires: 04/09/2021
Solution Control Lot: 19110 Bottle No.: 0648

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:59S		
Control 1 EC	0.095%	03:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	03:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:00S		
Control 2 EC	0.095%	04:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	04:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:01S		
Control 3 EC	0.094%	04:02S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.097%	04:02S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	04:03S		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GILROY

First Name: PATRICK

MI: W

Signature: 

Badge No.: 115

Date: 12/11/2019