

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXB-0076
Location:	SPRING LAKE POLICE		
Calibration File No.:	00506	Calib. Date:	12/03/2019
Certification File No.:	00507	Cert. Date:	12/03/2019
Linearity File No.:	00508	Lin. Date:	12/03/2019
Solution File No.:	00524	Soln. Date:	03/01/2020
Sequential File No.:	00524	File Date:	03/01/2020
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDYB S3-0001
Solution Control Lot:	19110	Expires:	04/09/2021
		Bottle No.:	0649

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:38S		
Control 1 EC	0.098%	17:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	17:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:39S		
Control 2 EC	0.097%	17:40S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	17:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:41S		
Control 3 EC	0.096%	17:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	17:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HEINE

First Name: JAMES

MI: E

Signature: PH Jenny Hei #121

Badge No.: 121

Date: 03/01/2020